## Fenwal International, Inc

Carretera Sanchez Km 18.5 Parque Industrial Itabo, Zona Franca Ind. de S.C. Haina, Dominican Republic

## INTERNATIONAL PRODUCT CERTIFICATION

ODE: R6R2004			BATCH: FA22K29019 '	
DESCRIPTION: COMPOFLEX 4F	- EMPTY - 150/150/1	50/150 - COL	IPLER - SEG. TUBE	
DATE OF MANUFACTURE: 11/29		EXPIRATION DATE: 2025-11		
FOR ENDOTOXIN TEST, G.M.P. AN	D CE			/ITH U.S. F.D.A., EDQM STANDAR
DATE BACTERIAL ENDOTOXIN	S TEST PASSED (C	HROMOGEN	IIC OR TURBIDIMETRIC	METHOD): 12/02/2022
_IMIT		:	0.250 <b>EU/ML</b>	
SAMPLE RESULT		:	<0.0100 EU/ML	
STERILIZATION METHOD: N.	/A GAMMA E-B	EAM EO	DATE STERILIZATION PI	ERFORMED: 12/12/2022
IINIMUM IRRADIATION DOSE:	PER SPEC:	14.2 <b>K</b> G	Y PER THIS LOT: 1	9.7 <b>KGY</b>
IAXIMUM IRRADIATION DOSE:	PER SPEC:	90.0 <b>KG</b>	Y PER THIS LOT: 4	6.5 <b>KGY</b>
		TEST OF	PRODUCT	
TEST		SPEC REFERENCE		RESULT
VISUAL INSPECTION		20-03-03-019		Passed ☐ Failed ☐
BACTERIAL ENDOTOXINS TEST		13-01-T		Passed ☐ Failed ☐
Functional Inspection, including leak test		20-03-03-019		Passed ☐ Failed ☐
FINAL FUNCTIONAL		20-03-03-019		Passed ☐ Failed ☐
OTHER TEST: N/A		N/A		Passed   Failed
SUPPLIER SOL	UTION BAG LO	OTS USED	IN CLOSED KITS (ONLY APP	PLY AMICUS APHERESIS KIT)
0 .9% NACL BAG ASSY W/OVER POUCH	ACD-A BAG ASSY W/OVER POUCH		N/A	N/A
N/A	N/A		N/A	N/A
N/A	N/A		N/A	N/A
BATCH RELEASE DATE: 01/31/2	2023		<b>QTY:</b> 6,	432 UNITS
PREPARED BY/ DATE: 47	01	3/-2023 NOT APPLIC	QUALITY APPROVER/ DATE:	347012 la Rose

REFERENCE DOCUMENT: SOP-DR0301004 Page 1 of 1 FORM NUMBER: DR0301004F1

FRESENIUS KABI Confidential REVISION: E