

**** **CERTIFICATE OF ANALYSIS** ****

Article:	Compoflex® 1F Empty - PD-V		
Article number:	T2131		
Batch number:	85WA02CC00	Quantity:	5550
Manufacturing date:	2024.01	Expiry date:	2026.12

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform

<u>Batch released by Quality Assurance.</u> Date: 18 Jan. 2024 Name: Wei Yuchan Signature: Wei Yuchan	<u>Manufactured by:</u> Fresenius Kabi AG Plant: FKGZ
--	---