



**FRESENIUS
KABI**

***** **CERTIFICATE OF ANALYSIS** *****

Article: CompoStop Flex2F PLT processing and leuko red.syst
Article number: FD51600
Batch number: 41UL07GA00 Quantity: 15 480
Manufacturing date: 3 November 2023 Expiry date: 2025-10

| Test | Limits | Results |
|-----------------------|-----------------------------------|----------------|
| Configuration | Conform product specification | Conform |
| Physical control | In accordance with specifications | Conform |
| Bacterial Endotoxines | < 20 EU/device | Conform |
| Sterility | Sterile (Ph.Eur.) | Conform |
| Used Filterbatch | 23PA001,23PA002 | |

| | |
|---|--|
| <u>Batch released by Quality Assurance.</u> | <u>Manufactured by:</u> |
| Date: 23.11.2023 | Fresenius Kabi AG |
| Name: Klímová Vladimíra | Plant: Fresenius Kabi Horatev CZ s.r.o. |
| Signature: <i>V. Klímová</i> | Hořátev 104, 289 13 Czech Republic |