



**FRESENIUS  
KABI**

\*\*\*\* **CERTIFICATE OF ANALYSIS** \*\*\*\*

Article: CompoFlex® 1F Empty - PD-V  
Article number: T2131  
Batch number: 85WB21CA00                      Quantity: 2764  
Manufacturing date: 2024.02                      Expiry date: 2027.01

| <b>Test</b>           | <b>Limits</b>                     | <b>Results</b> |
|-----------------------|-----------------------------------|----------------|
| Configuration         | Conform product specification     | Conform        |
| Physical control      | In accordance with specifications | Conform        |
| Bacterial Endotoxines | < 20 EU/device                    | Conform        |
| Sterility             | Sterile (Ph.Eur.)                 | Conform        |

|   |   |
|---|---|
| <u>Batch released by Quality Assurance.</u><br>Date: 22 Mar 2024<br>Name: Xu Xiaoying<br>Signature: | <u>Manufactured by:</u><br>Fresenius Kabi AG<br>Plant: FKGZ |
|---|---|

Form.no. 17-4228/2