LOVO X6R4940A Product Performance Report




Important: If reaction or injury has occurred call Fresenius Kabi Post-Market Quality Assurance at 1-800-933-6925.

		Software Version:	UDI No.:	
When Was the Proble ☐ Before Use ☐ Kit		t Check	☐ During Procedure	☐ After Procedure
Tubing: ☐ Flattened Separation Device: ☐ Pressure Pod: ☐ Cra Container: ☐ Leakir	ging Open □ Mispacked □ d □ Kinked □ Hole □ Cu □ Cracked □ Leaking Fluid acked □ Leaking Fluid □ Po	oor Fit on Pressure Sensor Port und Container Port □ Discolored	□ Discolored	
Additional Problem	Description / Explanation			

LOVO Cell Processing Disposable Kit with Dual Kit Check Vents (A)

X6R4940A LOVO Disposable Kit



Note:Kit layout subject to minor deviations from actual kits.

Please circle specific components on the diagram where incident occurred

Picture available for evaluation? Yes \square No \square

If a picture is available, please e-mail a clear picture along with this report to mdpmqa.usa@fresenius-kabi.com

Continued on Next Page

Please answer the following questions: 1. Was there any adverse event or injury? Yes □ No□ 2. Was the procedure successfully completed? Yes □ No□ 3. If no, was the procedure stopped due to a soft goods incident 4. Was product lost? Yes □ No□ N/A□ 5. Did the procedure involve clinical or patient material? Yes □ Check box if you do NOT wish to receive response letters. □	? Yes □ No □ N/A □	
	E-mail address for letter recipient (if applicable)	
Kit Return to Fresenius Kabi 1. Sample available for evaluation? Yes □ No □	Customer Information (please print) The following information is required to receive a credit	
2. Return label needed? Yes □ No □3. Sample return box needed? Yes □ No □	Facility Name:	
Center Authorized Signature/Date:	Account Number (if known):	
Fax this report to 1-888-858-2983 or E-mail this report to mdpmqa.usa@fresenius-kabi.com and include a copy of this form when returning a kit.		