

## CERTIFICATE OF ANALYSIS

Article:

Compoflex® 1F Empty - PD-V

Article number:

T2131

Batch number:

85WB02CA00

Quantity: 5525

Manufacturing date:

2024.02

Expiry date: 2027.01

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform

Batch released by Quality Assurance.

Date: 18 Mar. 2024

Name: Xu Xiaoying

Signature: Yu Xivo)il

Manufactured by:

Fresenius Kabi AG

Plant: FKGZ

Form.no. 17-4228/2